



Statement of Accuracy
Dissolved Oxygen Vacu-vials® Kit, Cat. No. K-7513

This statement of accuracy is based on laboratory tests performed under ideal testing conditions using standards of known concentration prepared in deionized water.

Instrument Platform	Test Range, ppm	Accuracy (expressed as ppm at concentrations across the test range)		
		2.0 ppm	4.0 ppm	11.0 ppm
Spectrophotometer	0 - 15.0	+/- 0.6	+/- 0.8	+/- 1.7
V-2000	0 - 15.0	+/- 0.6	+/- 0.8	+/- 1.7
V-3000	0 - 15.0	+/- 0.6	+/- 0.8	+/- 2.2
I-2002 SAM	0 - 15.0	+/- 0.6	+/- 0.8	+/- 1.7

Updated 12/01/2020