



**Statement of Accuracy**  
**Nitrate (as N) Vacu-vials® Kit, Cat. No. K-6983**

This statement of accuracy is based on laboratory tests performed under ideal testing conditions using standards of known concentration prepared in deionized water.

Instrument Platform	Test Range, ppm	Accuracy		
		0 ppm	1.0 ppm	11.25 ppm
Spectrophotometer	0 - 15.0	≤ 0.40	+/- 0.3	+/- 1.0
V-3000	0 - 15.0	≤ 0.40	+/- 0.3	+/- 1.0

Created 6/13/2025